

03 June 2019

W. Shearon

**SETTINGS FOR ETHIDIUM BROMIDE GEL VISUALIZATION IN THE COMMUNITY ROOM**

**COMPUTER SCREEN SETTINGS**

Contrast: 31  
Brightness: 5

**MOUNTED CAMERA SETTINGS**

Flip the “Trans” switch only  
Camera brightness all the way up

**GEL TRAY SURFACE SETTINGS**

UV setting: High  
Wavelength: 312 nm  
White light dial set to 5.5