



Heart Rate		Temperature	
GS	Spree	GS	Spree
82	82	94.75	95
83	82	96.1	95
83	82	96.48	95
83	82	96.48	95
81	82	96.48	95
69	110	95.03	95
94	112	95.57	95
97	96	96.1	95
104	98	96.1	95
103	116	96.1	95
107	107	95.5	95
99	112	95.5	95
96	110	97.7	95
123	120	96.93	95
114	114	95.4	95
98	127	96.1	95
103	133	96.1	95
99	100	97	95
87	90	96.08	95
95	100	97.33	95
93	90	96.8	95
105	102	97.18	95
79	75	96.3	95
75	66	95.2	95
73	80	97.7	95
71	71	96.4	95
83	73	96.1	95
101	165	97.3	95
109	142	98.7	95
100	142	98.9	95
127	126	97.4	95
117	130	96.9	95
127	123	97	95
103	127	96.8	95
128	140	95.7	95
110	129	97.1	95
128	132	96.9	95
108	133	98.8	95
155	157	97.2	95
123	118	96.8	95
115	109	98.6	95
84	110	98.5	95
112	104	98.6	95
102	114	98.5	95

Temperature GS
Temperature Spree
Heart Rate GS
Heart Rate Spree



82	70	94.2	95
79	89	95.2	95
75	91	96.8	95
90	96	97.5	95
92	97	97.3	95
105	92	97.1	95
88	96	96.7	95
102	87	97.5	95
108	95	98.3	95
101	95	97.2	95
95	91	97.8	95
99	96	98.4	95
105	97	98.6	95
102	97	97.1	95
105	95	99.2	95
99	96	97.6	95
100	87	97.9	95
98	91	98.3	95
85	91	97.4	95
82	86	98.6	95
85	79	93.4	95
80	72	95.6	95
79	78	97.1	97
81	81	97.8	95
100	96	97.1	97
78	88	97.6	97
76	94	97.6	97
124	90	98	97
125	103	96.1	95
79	99	94.8	95
59	104	94.4	95
111	119	94.7	95
108	108	94.2	95
106	106	95.5	95
100	102	95.2	95
101	99	95.6	95
110	100	95.7	95
99	115	96	95
104	100	96.5	95
99	89	97	95
83	87	96.6	95
100	86	96.6	95
76	93	97.1	95
65	82	98.04	95
90	85	98.34	97
95	86	98.04	97
87	85	98.5	97

94	80	97.5	97
95	92	97.86	95
103	99	96.76	95
107	118	98.34	95
101	109	98.5	95
99	110	98.04	95
98	110	96.08	97
130	148	96.63	95
108	120	98.08	95
112	118	97.9	95
114	117	97.33	95
111	114	98.17	95
108	116	98.75	95
115	114	98.48	95
95	104	98.62	95
101	97	98.5	95
94	90	97.6	95
96	97	97.6	95
122	125	94.82	95
105	115	94.28	95
157	157	96.8	97
118	130	97.88	95
124	126	98.24	97
97	116	96.98	95
111	115	97.34	95
119	194	94.5	95
142	150	96.44	95
189	179	96.44	95
112	147	95.9	95
136	153	95.08	95
170	168	96.98	95
157	157	96.44	97
174	175	97.16	95
130	128	96.62	97
117	162	97.52	97
112	134	96.44	97
119	133	98.08	97
106	127	97.7	97
114	115	98.06	97
122	123	97.57	97
81	81	96.6	95
74	70	96	95
84	72	96.5	95
90	77	96.8	95
85	73	96.9	95
79	83	95	95
83	89	96.5	95

86	90	96	95
81	92	97.7	95
66	97	97.4	97
96	83	97.5	97
96	78	97	97
98	83	97.2	97
83	81	97.3	97
91	88	97.5	97
98	93	97.7	97
100	77	95.7	97
90	92	96.3	95
96	92	96	95
90	97	97	95
90	94	96.3	95
97	98	97	95
79	81	97.3	95
80	81	96.6	97
84	82	96.6	97
78	80	96.3	95
79	80	96.3	95
94	93	95.2	95
96	94	94.6	95
98	96	94	95
102	95	95.4	95
104	95	95.2	95
101	96	95.3	95
99	97	95.2	95
100	98	94.1	95
100	98	93.9	95
102	97	94.2	95
101	96	94.5	95
102	97	94.6	95
77	80	92.1	95
81	81	93.1	95
93	79	96.5	95
81	81	97.1	95
73	73	96.5	95
85	83	97.3	97
78	67	97.6	97
72	82	97.3	97
97	68	97.1	97
69	68	97.9	97
122	131	97.3	97
119	91	96.3	96
144	81	96.8	96
82	81	96.2	96
112	83	96.5	96

79	75	96	96
86	79	96	96
125	82	96.9	97
75	75	96.5	96
105	100	96.9	96
120	85	96.4	96
121	82	96.5	96
63	64	97.2	96
69	71	97.4	96
62	60	96.9	96
67	67	96.7	96
54	72	96.4	97
91	0	97.9	95
88	84	98.7	95
82	86	98.7	95
94	89	98.7	95
80	102	98.1	95
119	124	97.9	95
0	126	97.4	95
123	88	96.9	95
119	118	98	95
136	118	97.7	95
126	107	97.4	95
0	113	97.9	95
105	104	97.2	95
123	120	96.6	95
104	110	96.4	95
108	112	97.7	95
122	112	97.9	95
106	108	97.5	95
85	91	98	95
83	87	98	95
98.2	80	79	95
80	81	98.5	95
69	70	97.5	95
74	70	97.6	95
73	70	97.7	95
76	71	97.8	95
73	81	97.8	97
83	73	97.6	95
80	69	97.2	95
62	95	95.8	95
107	107	96.1	95
132	131	96.7	95
100	97	96	95
104	101	97.7	95
106	107	97.8	95

109	107	98	95
96	91	97.5	95
82	82	97.7	95
82	87	97.8	97
87	87	97.1	95
61	58	97.1	95
62	61	97.1	97
64	61	97.2	97
60	60	97.2	97
66	59	97.1	97
75	73	96.7	97
70	60	96.8	95
67	68	94.2	95
127	126	95	95
112	116	96.1	95
99	110	95.9	95
81	103	95.8	95
79	78	96.3	95
84	82	96.5	95
79	92	95.2	95
140	97	95.1	95
77	72	95.4	95
86	75	96.1	95
70	70	96.8	95
79	69	97	95
68	69	97.1	95
74	70	96	95
76	71	96	97
70	68	96	97
71	70	96.1	97
74	71	96.02	97
72	75	90.9	95
130	132	92.5	95
126	122	93.4	95
111	114	94	95
124	108	92.2	95
131	125	93	95
118	127	92.6	95
126	118	90.08	95
131	123	90.1	95
116	110	90.5	95
110	106	90.6	95
110	106	98.6	95
110	106	90.88	95
107	103	93.7	95
107	103	94.48	97
107	103	94.1	97

	107	103	96.08	97
	107	103	98.2	95
	89	97	98.2	97
	93	91	98.7	97
	79	86	98.6	97
	92	86	98.7	97
	100	90	94.5	95
	81	49	97.5	95
	126	160	97.5	95
	128	192	97.9	95
	193	127	95.1	95
	130	127	97.5	95
	130	127	97.7	95
	100	121	97.4	95
	126	118	98.3	95
	126	126	98.4	95
	121	128	98.4	95
	125	127	98.4	95
	124	118	98.2	95
	96	96	98.2	95
	94	92	98.2	95
	91	84	98.2	95
	93	54	98.2	95
	98	102	98.5	97
Average	98.08977	98.94079	98.5	97
Standard Deviation	23.03054	24.83749	98.4	97
Count	303	304	98.6	97
Standard Error	1.336626	1.386403	98.3	97
			97.5	97
			88.2	97
			98.5	97
			98.5	97
			98.2	97
			98	97
			97.9	98
			97.5	97
			98.8	97
			97.6	97
			97	97
			98.2	97
			97.1	97
			97.8	97
			98.1	97
			98.5	97
			98.3	98
Average			96.6527	95.53086
Standard Deviation			1.918463	0.870378

Count

326

324

Standard Error


0.10907225

0.048067


Average	Standard Deviation
96.6527	1.918463
95.53086	0.870378
98.08977	23.03054
98.94079	24.83749

Post-hoc Tests	t-test value
Heart Rate: GS vs Spree	0.6574608
Temperature: GS vs Spree	3.366E-20

Gold Standand VS Spree



Gold Standand VS Spree



Corrected p-value to achieve significance Significant?

0.025 no

0.025 yes