

pipette	water mass	calculated volume	theoret volu	% error
P-200	0.20133	0.202041185	0.2	1.02059249
	0.4345	0.436034843	0.4	9.00871066
	0.603335	0.605466241	0.6	0.91104019
	0.80627	0.809118096	0.8	1.13976196
	1.00982	1.013387123	1	1.33871227
	1.21356	1.217846821	1.2	1.48723507

glass mass	13.1426
density@ 27.8	0.99648